

ENTREPRENEURSHIP DEVELOPMENT CENTER
Clarification on Tender Number- EDC/NIT/2019/005
Item Description – Procurement of Surface Plasmon Resonance Instrument

A press tender notice was issued on 8th November 2019 for the procurement of Surface Plasmon Resonance Instrument under two-bid system.

A pre-bid meeting was held on 19th November 2019 at 11:00 am and the minutes of the meeting is as under:

- 1) Prebid meeting was attended by representatives from ForteBio Biologics - Molecular Devices & GE Healthcare.
- 2) Prior & during the meeting written queries were received from representatives. Our replies to the same are as summarized below:

Specification Ref.	Query	Our reply:
Sr. No. 1,4,5	Please Change to: 4 - 40°C for both analysis and sample chambers temperature.	No amendment
Sr. No. 3	System should be able to study mM (millimolar) to pM (picomolar) range of affinities.	No amendment
Sr. No. 7	3 or More channels.	No amendment
Sr. No. 11	One Buffer system or better.	No amendment
Warranty	We would like to request for making provision to quote the system with one year of standard warranty and remaining four years of AMC or CMC as optional items. Being the manufacturers of SPR systems, we feel that keeping five years of CMC will increase the cost due to operational reasons associated with CMC.	Will be amended. Please refer to Point 3 of the Minutes below. Visit our website www.venturecenter.co.in to see the amendment document.
Sr. No. 23 & 24	Mentioned sensor surfaces are reusable surfaces, where in one surface can be used for more than 500 cycles. Keeping 10 units of each of these types will be more because of	No amendment

	the shelf life of these sensor surface-chemistries (because they will most probably not be used within stipulated expiry date).	
--	--	--

3) Clause 3.10.1 is amended as follows:

Existing Nomenclature: The items covered by the schedule shall carry minimum five years of comprehensive warranty from the date of acceptance of the equipment by EDC.

Amended Nomenclature: The items covered by the schedule shall carry two years of comprehensive warranty from the date of acceptance of the equipment by EDC and Annual Maintenance Contract in years 3,4 & 5.

4) Clause 3.11.2 is amended as follows;

Existing Nomenclature: One week advanced training for two users nominated by EDC on operation, method development, maintenance, software, data interpretation(qualitative and quantitative, unknown identification), etc. at the manufacturer's international application laboratory

Amended Nomenclature: One week advanced training for two users nominated by EDC on operation, method development, maintenance, software, data interpretation(qualitative and quantitative, unknown identification), etc. at the manufacturer's application laboratory in India.

Kindly refer to the Amendment Document on our website: www.venturecenter.co.in for more details.

5) In general the bidders expressed their satisfaction about the details and Specifications given in the bid document.

The terms and conditions of the notice issued on our EDC website www.venturecenter.co.in will remain unchanged.

21.11.19

General Manager (EDC)