

ENTREPRENEURSHIP DEVELOPMENT CENTER

Amendments to Tender Number- EDC/Analytical Services/2025-26/004

Item Description: Procurement of CHNSO Analyzer

A press tender notice was issued on 13th January 2026 for the procurement of CHNSO analyzer under the two-bid system. The following amendments were recommended by the technical committee.

Specification Reference	Existing Nomenclature	Amended Nomenclature
Ch 4 Sr. No. 4	Supplier must clearly specify their configuration approach and provide either : Dual furnace system	Supplier must clearly specify their configuration approach and provide with a Dual furnace system
Ch 4 Sr. No.6	Should be equipped with automated injection with consistent sample delivery.	Should be equipped with automated injection with consistent sample delivery. The autosampler should be electrical/electromechanical with fully software controlled operation
Ch 4 Sr. No. 8	Furnace temperature range: 950°C-1100	Furnace temperature should be capable of reaching 1100°C during combustion with an integrated or separate reduction arrangement
Ch 4 Sr. No. 11	Standard deviation : 0.01% of absolute (to be supported with evidence document)	Standard deviation : 0.1% of absolute (to be supported with evidence document)
Ch 4, Sr. No. 14	Nitrogen or compressed air for pneumatic auto sampler requirements	Nitrogen or compressed air for auto sampler requirements
Ch 4, Miscellaneous	Nitrogen or compressed air for pneumatic auto sampler requirements	Nitrogen or compressed air for auto sampler requirements

The terms and conditions of the notice issued on our EDC website
<https://www.venturecenter.co.in/tenders/> will remain unchanged.

27.01.2026

Manisha
 Chief Operating Officer (EDC)
 (Manisha Premnath)